

बीरबल साहनी पुरावनस्पतिविज्ञान संस्थान, लखनऊ
BIRBAL SAHNI INSTITUTE OF PALAEOBOTANY, LUCKNOW

टेलीग्राम पेलियोबॉटनी

53, विश्व विद्यालय मार्ग

दूरभाष 2740008, 2740011

लखनऊ - 226007

फैक्स ९१-91-522-2740485, 2740098

बी सा पु सं /III/भंडार एवं क्रय/ A.S C-957

Date: 12.11.2015

Speed Post

13

Laboratory Oven

Attach Authorization or ISO Certificate alongwith quote separately

Subject: -----

Dear Sirs,

This Institute intend to purchase item mentioned above. Sealed quotations are invited so as reach this office on or before 5:00P.M. on 27.11.2015 duly superscribed by "Oven" due to open in presence of vendors at a later date to be notified.

While submitting quotation please note that:

1. The material may either of indigenous manufacturer or of foreign make, available from ready stock. Any offer to supply on forward Delivery Basis under suppliers own quota license will also be considered.
2. The price quoted should be F.O.R. Destination.
3. Your rates should include packing, insurance and forwarding charges.
4. The rates of Sales Tax should be clearly indicated wherever chargeable. The tendered should also indicate Central/Sales Tax Registration Number and date in this quotation.
5. Specific mention should be made whether the offer is for supplies available ex-stock. In case the officer is on Forward Delivery basis, firm delivery period must be indicated.
6. The cover should be sealed and superscribed "Quotation for "Oven" must be written on envelope in the name of "Director, Birbal Sahni Institute of Palaeobotany, Lucknow". The quotations not complying the procedure will be rejected.
7. Payments will be made by crossed cheque on The Indian Overseas Bank, Lucknow only after receipt and acceptance of supply and installation/ if required satisfactory.
8. The acceptance of the quotation will rest with the Director who does not bind himself to accept the lowest quotation and reserves the right to himself to reject or partially accept any or all the quotation received without assigning any reasons.
9. The quotations are liable to be cancelled if any of the above mentioned conditions are not complied with.

Yours sincerely,

(Ram Shukla)
Registrar

✓ Convener, Website Committee, BSIP, Lucknow

Specification for Laboratory Oven

The laboratory oven should be easy to install, which should have environmentally friendly design with cost savings, reducing energy consumption and have a low cost of operation/ownership be easy to use. Should have normal warranty.

It should be large, easy to view vacuum fluorescent display, easy setting with intuitive interface and easy cleaning with rounded corners.

Chamber Volume	400-500 Liters
Convection	Gravity Convection
Temperature Range	50°C to 250°C
Temp Uniformity	±3°C at 150°C
Temp Sensitivity	±0.5°C
Chamber Dimensions	Minimum 51"x 21"x 23"
Ext Dimensions	Minimum 60"x 30"x 30"
Weight	Less than 500 lbs
Power Output	Preferable 2640 Watts / 11 Amps
Accessories	5 Shelves
Electrical	208-240V 50/60Hz
Low noise during operation	

- Standard wheel (which can be locked) for easy movement.
- Instruction manual should be provided. 12 months Warranty with documents should be provided.
- Supporting documents wherever required and original catalogue should be provided to validate specifications.